

EU DECLARATION OF CONFORMITY

| | |
|---|--|
| DOC No. | DOC-MYMEDİKAL-SRT-002 |
| EC Certificate | Not applicable (Self- declared) |
| Manufacturer | MY TICARET VE MEDİKAL A.S. |
| Manufacturer Address | Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey |
| Single Registration Number (SRN) | TR-MF-000018372 |
| Brand | Mumu |
| Product Description | Powdered Latex Examination Gloves |
| Intended Purpose | A patient examination glove is a medical device intended for a medical purpose that is worn on the examiner's hand or finger to prevent contamination between the patient and examiner. Examination glove is intended for medical activities except surgery. |
| Basic UDI-DI | 868227994LP5H |
| Size | XS, S, M, L, XL |
| EAN Code | 8683020022980, 8681695367078, 8681695367085, 8681695367092, 8683020021860 |
| Country of Origin | Thailand |
| European Medical Device Nomenclature (EMDN) | T010201 (Examination/Treatment Gloves, Latex) |
| Global Medical Device Nomenclature (GMDN) | 47173 (Latex examination/treatment glove, powdered) |
| Product Catalogue/Reference Number | MLP01-XS, MLP02-S, MLP03-M, MLP04-L, MLP05-XL |
| Product Group Reference Number | LX01 |
| Conformity Assessment Route (MDR): | Annex II and Annex III according to EU 2017/745 |
| Classification & Rule (MDR) | Class I, Rule 5 transient use according to Annex VIII |
| Device Classification (PPER) | Category III |
| EU Type-Examination Certificate (PPER) | 2777/10468-05/E05-01 |
| Notified Body (PPER) | EU-Type Examination and Ongoing Conformity by Notified Body SATRA TECHNOLOGY EUROPE LTD Bracetown Business Park, Clonee, D15YN2P, Ireland [CE 2777] |



MY Medikal

MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

| Applicable Standards | No. | Regulation/ Standard Number | Regulation/ Standard Name |
|----------------------|--------------------|---|---------------------------|
| | 1 | MDR (EU) 2017/745 | Medical Device Regulation |
| 2 | PPE (EU) 2016/425 | Personal Protective Equipment Regulation | |
| 3 | ISO 13485: 2016 | Medical devices - Quality management systems - Requirements for regulatory purposes | |
| 4 | ISO 9001: 2015 | Quality management systems – requirements | |
| 5 | ISO 14971: 2019 | Medical devices - application of risk management to medical devices | |
| 6 | EN 455-1: 2020 | Requirements and testing for freedom from holes | |
| 7 | EN 455-2: 2015 | Requirements and testing for physical properties | |
| 8 | EN 455-3: 2015 | Requirements and testing for biological evaluation | |
| 9 | EN 455-4: 2009 | Requirements and testing for shelf-life determination | |
| 10 | ISO 10993-10: 2010 | Biological evaluation of medical devices –Part 10: Test for irritation and skin sensitization | |
| 11 | ISO 20417: 2021 | Medical Devices- Information to be supplied by the manufacturer | |
| 12 | ISO 15223-1: 2021 | ISO 15223-1 Symbols to be used with information to be supplied by the manufacturer | |
| 13 | EN ISO 374-1: 2016 | Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks | |
| 14 | EN ISO 374-2:2014 | Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration | |
| 15 | EN ISO 374-4:2013 | Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals | |
| 16 | EN ISO 374-5: 2016 | Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology | |

| | | | |
|--|----|------------------|--|
| | | | and performance requirements for micro-organisms risks |
| | 17 | EN 16523-1: 2015 | Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact |

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentation is retained under the premise of the manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System.
- Is following the EU-Type Examination with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under the surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance with Annex IV of Medical Device Regulation (EU) 2017/745.

Authorized Signatory:

Approver : MURAT YILDIZ
Title : General Manager/CEO
Signature : 
Approval Date : 23 Sep 2024
Place of Approval : Istanbul, Turkey

MY TICARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mahallesi General Şükrü Koraltı Cad
No:33 Arnavutköy/İSTANBUL
Büyükdere Şişli V.D.826 040 4605
Tel:02124382064 Fax:0212 438 20 65
www.mymedikal.com

