



MY Medikal

MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr

EU DECLARATION OF CONFORMITY

| EC Certificate | Not applicable (Self- declared) | | | | | | | | | | | | |
|---|--|--|-----------------------------|---------------------------|---|-------------------|---------------------------|---|-------------------|--|---|-----------------|--|
| Manufacturer | MY TICARET VE MEDİKAL A.S. | | | | | | | | | | | | |
| Manufacturer Address | Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey | | | | | | | | | | | | |
| Single Registration Number (SRN) | TR-MF-000018372 | | | | | | | | | | | | |
| Brand | E-Care | | | | | | | | | | | | |
| Product Description | Nitrile Powder Free Examination and Protective Gloves | | | | | | | | | | | | |
| Intended Purpose | A patient examination glove is a medical device intended for a medical purpose that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Examination glove is intended for medical activities except surgery. | | | | | | | | | | | | |
| Basic UDI-DI | 868302002NPVQ | | | | | | | | | | | | |
| Size | XS, S, M, L, XL | | | | | | | | | | | | |
| European Medical Device Nomenclature (EMDN) | T01020204 (Examination / Treatment Gloves, Nitrile) | | | | | | | | | | | | |
| Global Medical Device Nomenclature (GMDN) | 56286 (Nitrile Examination/Treatment glove, non-powdered, non-sterile) | | | | | | | | | | | | |
| Product Catalogue Number | ENVXS00, ENVS01, ENVM02, ENVL03, ENVXL04 | | | | | | | | | | | | |
| Conformity Assessment Route (MDR): | Annex II and Annex III according to EU 2017/745 | | | | | | | | | | | | |
| Classification & Rule (MDR) | Class I, Rule 1 & Rule 5 according to Annex VIII | | | | | | | | | | | | |
| Device Classification (PPER) | Category III | | | | | | | | | | | | |
| Product Group Reference Number | SNBE20013, SNBE20014, SNBE20015, SNBE20016, SNBE20017 | | | | | | | | | | | | |
| EU Type-Examination Certificate (PPER) | 2777/14815-03/E00-00 | | | | | | | | | | | | |
| Notified Body Number (PPER) | 2777 | | | | | | | | | | | | |
| EU Type- Examination Certificate Issued by (PPER) | SATRA Technology Europe Limited Bracetown Business Park, Clonee, D15YN2P, Ireland | | | | | | | | | | | | |
| Applicable Standards | <table border="1"><thead><tr><th>No.</th><th>Regulation/ Standard Number</th><th>Regulation/ Standard Name</th></tr></thead><tbody><tr><td>1</td><td>MDR (EU) 2017/745</td><td>Medical Device Regulation</td></tr><tr><td>2</td><td>PPE (EU) 2016/425</td><td>Personal Protective Equipment Regulation</td></tr><tr><td>3</td><td>ISO 13485: 2016</td><td>Medical devices - Quality management systems -</td></tr></tbody></table> | No. | Regulation/ Standard Number | Regulation/ Standard Name | 1 | MDR (EU) 2017/745 | Medical Device Regulation | 2 | PPE (EU) 2016/425 | Personal Protective Equipment Regulation | 3 | ISO 13485: 2016 | Medical devices - Quality management systems - |
| No. | Regulation/ Standard Number | Regulation/ Standard Name | | | | | | | | | | | |
| 1 | MDR (EU) 2017/745 | Medical Device Regulation | | | | | | | | | | | |
| 2 | PPE (EU) 2016/425 | Personal Protective Equipment Regulation | | | | | | | | | | | |
| 3 | ISO 13485: 2016 | Medical devices - Quality management systems - | | | | | | | | | | | |



MY Medikal

MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

| | | |
|----|-----------------------------|--|
| | | Requirements for regulatory purposes |
| 4 | ISO 9001: 2015 | Quality management systems – requirements |
| 5 | ISO 14971: 2019 | Medical devices - application of risk management to medical devices |
| 6 | EN 455-1: 2020 | Requirements and testing for freedom from holes |
| 7 | EN 455-2: 2015 | Requirements and testing for physical properties |
| 8 | EN 455-3: 2015 | Requirements and testing for biological evaluation |
| 9 | EN 455-4: 2009 | Requirements and testing for shelf-life determination |
| 10 | ISO 10993-1: 2018 | Biological evaluation of medical devices –Part 1: Evaluation and testing within a risk management process |
| 11 | ISO 10993-10: 2010 | Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization |
| 12 | ISO 10993-11: 2017 | Biological evaluation of medical devices — Part 11: Tests for systemic toxicity |
| 13 | EN 1041: 2008+A1: 2013 | Information supplied by the manufacturer of medical devices |
| 14 | ISO 15223-1: 2021 | Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements |
| 15 | EN ISO 374-1: 2016+A1: 2018 | Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks |
| 16 | EN ISO 374-2: 2019 | Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration |
| 17 | EN ISO 374-4: 2019 | Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals |
| 18 | EN ISO 374-5: 2016 | Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks |



MY Medikal

MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

| | | | |
|--|----|------------------------------|--|
| | 19 | EN 16523-1: 2015+A1: 2018 | Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact |
| | 20 | ASTM D 6978-05:2019 | Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs |
| | 21 | ASTMF1671/F1671-13 | Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System |

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentations are retained under the premise of manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System:

Verification Certificates: EN ISO 13486:2016 Quality Management System
Certificate No: ISO 02 836 1179

EN ISO 9001:2015 Quality Management System
Certificate No.: ISO 01 940 1179

- Is following to the EU-Type Examination and conformity with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance to Annex IV of Medical Device Regulation (EU) 2017/745.



MY Medikal


MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: www.mymedikal.com.tr.

Authorized Signatory:

Approver : MURAT YILDIZ
Title : General Manager/CEO
Signature : 
Approval Date : 25.10.2022
Place of Approval : Istanbul, Turkey

MY TICARET VE
MEDİKAL ANONİM ŞİRKETİ
Ömerli Mah. General Şükrü Koraltı Cad
No:33 Arnavutköy/İSTANBUL
Büyükdere Şişli V.D.626 040 4605
Tel:0212 438 20 64 Fax:0212 438 20 65
www.mymedikal.com

CE