MY TICARET VE MEDIKAL A.S.
Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey
Tel: +902124382064 Fax: +902124382065
Website: www.mymedikal.com.tr.

EU DECLARATION OF CONFORMITY

EC Certificate	Not applicable (Self- declared)
Manufacturer	MY TICARET VE MEDIKAL A.S.
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey
Single Registration Number (SRN)	TR-MF-000018372
Brand	Mumu Care
Product Description	Nitrile Powder Free Examination and Protective Gloves
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Examination glove is intended for medical activities except surgery.
Basic UDI-DI	868302002NPVQ
Size	XS, S, M, L, XL
European Medical Device Nomenclature (EMDN)	T01020204 (Examination / Treatment Gloves, Nitrile)
Global Medical Device Nomenclature (GMDN)	56286 (Nitrile Examination/Treatment glove, non-powdered, non-sterile)
Product Catalogue Number	MCN01-XS, MCN02-S, MCN03-M, MCN04-L, MCN05-XL
Conformity Assessment Route (MDR):	Annex II and Annex III according to EU 2017/745
Classification & Rule (MDR)	Class I, Rule 1 & Rule 5 according to Annex VIII
Device Classification (PPER)	Category III
Product Group Reference Number	SNBE20013-XS, SNBE20014-S, SNBE20015-M, SNBE20016-L, SNBE20017-XL
EU Type-Examination Certificate (PPER)	2777/14815-03/E00-00
Notified Body Number (PPER)	2777
EU Type- Examination Certificate Issued by (PPER)	SATRA Technology Europe Limited Bracetown Business Park, Clonee, D15YN2P, Ireland
Applicable Standards	

No.	Regulation/ Standard Number	Regulation/ Standard Name
1	MDR (EU) 2017/745	Medical Device Regulation
2	PPE (EU) 2016/425	Personal Protective Equipment Regulation
3	ISO 13485: 2016	Medical devices - Quality management systems -

			Requirements for regulatory purposes
	4	ISO 9001: 2015	Quality management systems – requirements
	5	ISO 14971: 2019	Medical devices - application of risk management to medical devices
	6	EN 455-1: 2020	Requirements and testing for freedom from holes
	7	EN 455-2: 2015	Requirements and testing for physical properties
	8	EN 455-3: 2015	Requirements and testing for biological evaluation
	9	EN 455-4: 2009	Requirements and testing for shelf-life determination
	10	ISO 10993-1: 2018	Biological evaluation of medical devices –Part 1: Evaluation and testing within a risk management process
	11	ISO 10993-10: 2010	Biological evaluation of medical devices — Part 10: Tests for irritation and skin sensitization
	12	ISO 10993-11: 2017	Biological evaluation of medical devices — Part 11: Tests for systemic toxicity
	13	EN 1041: 2008+A1: 2013	Information supplied by the manufacturer of medical devices
	14	ISO 15223-1: 2021	Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements
	15	EN ISO 374-1: 2016+A1: 2018	Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks
	16	EN ISO 374-2: 2019	Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration
	17	EN ISO 374-4: 2019	Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals
	18	EN ISO 374-5: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks

MY TICARET VE MEDIKAL A.S.
Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey
Tel: +902124382064 Fax: +902124382065
Website: www.mymedikal.com.tr.

	19	EN 16523-1: 2015+A1: 2018	Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact
	20	ASTM D 6978-05:2019	Standard Practice for Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs
	21	ASTMF1671/F1671-13	Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood-Borne Pathogens Using Phi-X174 Bacteriophage Penetration as a Test System

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:

- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentations are retained under the premise of manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System:

Verification Certificates: EN ISO 13486:2016 Quality Management System
Certificate No: ISO 02 836 1179

EN ISO 9001:2015 Quality Management System
Certificate No.: ISO 01 940 1179

- Is following to the EU-Type Examination and conformity with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance to Annex IV of Medical Device Regulation (EU) 2017/745.

MY TICARET VE MEDIKAL A.S.
Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey
Tel: +902124382064 Fax: +902124382065
Website: www.mymedikal.com.tr.

Authorized Signatory:

Approver : MURAT YILDIZ

Title : General Manager/CEO

Signature :

Approval Date : 25.10.2022

Place of Approval : Istanbul, Turkey