



MY Medikal

## MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: [www.mymedikal.com.tr](http://www.mymedikal.com.tr)

### EU DECLARATION OF CONFORMITY

EC Certificate	Not applicable (Self- declared)												
Manufacturer	MY TICARET VE MEDİKAL A.S.												
Manufacturer Address	Ömerli mah General Şükrü Koraltı Cd no:33, 34555 Arnavutkoy/Istanbul, Turkey												
Single Registration Number (SRN)	TR-MF-000018372												
Brand	Mumu Plus+												
Product Description	Powderfree Latex Examination Gloves												
Intended Purpose	A patient examination glove is a medical device intended for a medical purpose that is worn on the examiners hand or finger to prevent contamination between patient and examiner. Examination glove is intended for medical activities except surgery.												
Basic UDI-DI	868227994LPFX3												
Size	XS, S, M, L, XL												
European Medical Device Nomenclature (EMDN)	T010201 (Examination/Treatment Gloves, Latex)												
Global Medical Device Nomenclature (GMDN)	47172 (Hevea-latex Examination/treatment glove, non-powdered, non-antimicrobial)												
Product Catalogue Number	SMLF00-XS; SMLF01-S; SMLF02-M; SMLF03-L; SMLF04-XL												
Conformity Assessment Route (MDR):	Annex II and Annex III according to EU 2017/745												
Classification & Rule (MDR)	Class I, Rule 5 transient use according to Annex VIII												
Device Classification (PPER)	Category III												
Product Group Reference Number	MTCLPF												
EU Type-Examination Certificate (PPER)	2777/12719-01/E05-02												
Notified Body Number (PPER)	2777												
EU Type- Examination Certificate Issued by (PPER)	SATRA Technology Europe Limited Bracetown Business Park, Clonee, D15YN2P, Ireland												
Applicable Standards	<table border="1"><thead><tr><th>No.</th><th>Regulation/ Standard Number</th><th>Regulation/ Standard Name</th></tr></thead><tbody><tr><td>1</td><td>MDR (EU) 2017/745</td><td>Medical Device Regulation</td></tr><tr><td>2</td><td>PPE (EU) 2016/425</td><td>Personal Protective Equipment Regulation</td></tr><tr><td>3</td><td>ISO 13485: 2016</td><td>Medical devices - Quality management</td></tr></tbody></table>	No.	Regulation/ Standard Number	Regulation/ Standard Name	1	MDR (EU) 2017/745	Medical Device Regulation	2	PPE (EU) 2016/425	Personal Protective Equipment Regulation	3	ISO 13485: 2016	Medical devices - Quality management
No.	Regulation/ Standard Number	Regulation/ Standard Name											
1	MDR (EU) 2017/745	Medical Device Regulation											
2	PPE (EU) 2016/425	Personal Protective Equipment Regulation											
3	ISO 13485: 2016	Medical devices - Quality management											



MY Medikal

## MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: [www.mymedikal.com.tr](http://www.mymedikal.com.tr)

		systems - Requirements for regulatory purposes
4	ISO 9001: 2015	Quality management systems – requirements
5	ISO 14971: 2019	Medical devices - application of risk management to medical devices
6	EN 455-1: 2020	Requirements and testing for freedom from holes
7	EN 455-2: 2015	Requirements and testing for physical properties
8	EN 455-3: 2015	Requirements and testing for biological evaluation
9	EN 455-4: 2009	Requirements and testing for shelf-life determination
10	ISO 10993-10: 2010	Biological evaluation of medical devices –Part 10: Test for irritation and skin sensitization
11	EN 1041: 2008+A1: 2013	Information supplied by the manufacturer of medical devices
12	ISO 15223-1: 2021	ISO 15223-1 Symbols to be used with information to be supplied by the manufacturer
13	EN ISO 374-1: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 1: Terminology and performance requirements for chemical risks
14	EN ISO 374-2:2014	Protective gloves against dangerous chemicals and micro-organisms - Part 2: Determination of resistance to penetration
15	EN ISO 374-4:2013	Protective gloves against chemicals and micro-organisms - Part 4: Determination of resistance to degradation by chemicals
16	EN ISO 374-5: 2016	Protective gloves against dangerous chemicals and micro-organisms - Part 5: Terminology and performance requirements for micro-organisms risks
17	EN 16523-1: 2015	Determination of material resistance to permeation by chemicals - Part 1: Permeation by liquid chemical under conditions of continuous contact
18	EN 420:2003+A1:2009	Protective gloves - General requirements and test methods



## MY TICARET VE MEDİKAL A.S.

Ömerli Mahallesi General Şükrü Koraltı Caddesi No:33 Arnavutkoy –Istanbul Turkey

Tel: +902124382064 Fax: +902124382065

Website: [www.mymedikal.com.tr](http://www.mymedikal.com.tr)

We, My Ticaret ve Medikal A.S. herewith declare that the above-mentioned device:


- Is in compliance with the General Safety Performance Requirement of the Medical Device Regulation (MDR) 2017/745. All supporting documentations are retained under the premise of manufacturer.
- The gloves are manufactured according to EN ISO 9001:2015 and EN ISO 13485:2016 Quality Management System:

Verification Certificates: EN ISO 13486:2016 Quality Management System  
Certificate No: ISO 02 836 1179

EN ISO 9001:2015 Quality Management System  
Certificate No.: ISO 01 940 1179

- Is following to the EU-Type Examination and conformity with the provisions of new PPE Regulations (EU) 2016/425 Category III of the notified body number 2777 by SATRA Technology Europe Ltd.
- Is in conformity to type based on quality assurance of the production process under surveillance of the notified body number 2777 by SATRA Technology Europe Ltd.
- This EU Declaration of Conformity is prepared in accordance to Annex IV of Medical Device Regulation (EU) 2017/745.

### Authorized Signatory:

Approver : MURAT YILDIZ  
Title : General Manager/CEO  
Signature :   
Approval Date : 25.10.2022  
Place of Approval : Istanbul, Turkey

MY TICARET VE  
MEDİKAL ANONİM ŞİRKETİ  
Ömerli Mahallesi General Şükrü Koraltı Cad  
No:33 Arnavutköy/İSTANBUL  
Büyükdere Şişli Ce V.D.826 040 4605  
Tel:0212 438 20 64 Fax:0212 438 20 65  
[www.mymedikal.com](http://www.mymedikal.com)

